



SPECIAL REPRINT, JIRD[®] No. 1, 2017

Inside this issue:

Clinical guidelines for using T3[®] Short Implants

By Francesco Amato, MD, DDS, PhD

Clinical case presentations featuring T3 Short Implants

short

Clinical guidelines for using T3® Short Implants

Francesco Amato, MD, DDS, PhD[†], Italy



Although longer implants for many years were thought to be safest, with at least 10mm of implant length considered to be the standard, anatomical limitations often prohibit placement of such implants without undertaking invasive and potentially risky surgical interventions. More recently, good results have been found for the use of short implants to rehabilitate posterior partial edentulism in atrophic maxillary and mandibular bone. To increase the likelihood of long-term success for short implants, this article presents guidelines, including recommendations for platform-switching, treatment staging, splinting, osteotomy preparation, and more. Two maxillary cases illustrating the use of short implants are also presented.

Key words: implants, edentulous, resorption, guidelines, T3 Short Implants

Introduction

The crown-to-root ratio for natural teeth is often viewed as an indicator of tooth prognosis, with a minimum 1:1 ratio recommended and 1:2 seen as the ideal.¹ The increased functional lever arm of an unfavorable crown-to-root ratio is considered a non-axial loading force.¹ When dental implants were first introduced, similar guidelines were adapted. It was assumed that longer implants would prove more advantageous in clinical use than shorter ones, due both to the more favorable crown-to-implant ratio² and the greater implant surface area available for osseointegration. Implant dimensions of 4 mm in diameter and at least 10 mm in length became the standard and were considered to be safest,³ with 10 to 12 mm of residual alveolar bone thought to be the minimum necessary to ensure predictable implant treatment.

In the posterior region, however, that amount of bone height is frequently unavailable,⁴⁻⁶ and the bone quality may be compromised. The presence of the maxillary sinus or inferior alveolar nerve also may limit the availability of bone in posterior sites⁷ (Figs. 1a-b, 2a-b). To overcome such limitations, surgical procedures such as sinus lifts, vertical bone augmentation, guided bone regeneration, alveolar nerve transposition, and placement of tilted implants were developed.⁸⁻¹⁰ But these surgical procedures are substantially invasive and pose risks of intra- and post-operative complications, infection, or graft resorption.¹¹ Bone-augmentation surgeries also increase the length and cost of treatment.




Fig. 1a



Fig. 1b

Figs. 1a, b. Radiograph and Cone Beam CT scan image showing minimal bone height under the maxillary sinus.




Fig. 2a



Fig. 2b

Figs. 2a, b. Radiograph and Cone Beam Scan image showing reduced bone height above the inferior alveolar nerve canal.



An alternative to these surgical procedures is to use short implants.^{12,13} The term “short implants” has been controversial, with studies and reviews lacking consensus about its definition.¹⁴ In 1991 8 and 9 mm length implants were introduced and defined as “short.” Since then some authors have defined short implants as being less than 7 mm long, while others have extended the definition to include all implants with lengths of up to 10 mm¹⁵ (Fig. 3).

Throughout the 1990s, higher failure rates for shorter implants were reported by a number of investigators.¹⁶⁻²³ However, more recent studies have found better results using short implants to rehabilitate posterior partial edentulism in very atrophic maxillary and mandibular bone. Renouard and Nisand in 2005 reported a 94.6% survival rate after 2 years of loading on short implants placed with high initial stability and good bone-to-implant contact.²⁴ In a six-year multicenter retrospective study, Misch et al in 2006 found a 98.9% survival rate for 745 7 mm and 9 mm long posterior implants.²⁵ A 2012 systematic review by Annibali et al that analyzed results of two randomized controlled trials and 14 observational studies and included a total of 6,193 short implants, found a cumulative survival rate of 99.1%, with a low incident of biological and biomechanical complications.²⁶ Another extensive review of 33 studies of short implants published between 1980 and 2004 found the overall success rate to be 95.2%.²⁷ While the authors found poor bone quality to be associated with short implant failures, they concluded that the use of implants 4 mm in diameter appeared to minimize failure in such situations.

Several explanations have been offered for the improvement in outcomes for short implants that has become apparent over time. Most importantly, newer

surface treatments and wider diameters of short implants in use today increase the bone-to-implant contact exponentially. Whereas early implants had smooth (machined/turned) surfaces, various techniques have since been introduced to alter the implant surface topography, including acid-etching, grit blasting, titanium plasma-spraying, and nanoparticle deposition. These techniques both roughen and increase the implant surface area,²⁸ and they also have been found to accelerate osseointegration.²⁹ Evaluating the effect of titanium surface topography on bone integration, Wennerberg and Albrektsson concluded that surface roughness influences bone response at the micrometer level.³⁰ Many studies have concluded that the advances in surface topography and chemistry have made short implant survival rates comparable to those of standard length implants.^{15,31-36}

While some studies have found that neither implant length nor width significantly affects short implant survival rates,^{37,38} Anitua et al showed that crestal bone resorption around short implants decreased with increased implant diameter and that using wider implants can reduce the maximum von Mises stress in bone by 20 to 30%.³⁹

Other reports of finite element analyses support the hypothesis that the use of shorter implants in appropriate clinical situations yields cumulative survival rates comparable to those reported for longer implants. Lum found that occlusal forces applied to implants were distributed primarily to the crestal bone, regardless of implant length.⁴⁰ Lum and Osier also reported that masticatory forces were well tolerated by the crestal bone, but parafunctional forces were not and should be attenuated.^{41,42} Holmgren et al⁴³ and Himmlova et al⁴⁴ demonstrated that force application resulted in greatest force concentration at the bone crest. Himmlova et al stated that while implant length had no effect on either the magnitude of peak stress or stress distribution to the supporting bone, implant diameter was more important for improved stress distribution. When Anitua et al in 2010 conducted a finite element analysis of the influence of implant length, diameter, and geometry on implant surface stress distribution, they found stresses to be localized on the first six implant threads, independent of the implant length, diameter, or macrogeometry.⁴⁵ They also reported that at a constant diameter, the maximum stress value observed in the first six threads was equal or even lower in shorter implants (8.5 mm) than in longer ones.



Fig. 4. T3® Short Implant and Final Shaping Drill
The last drill used before the insertion of a T3 Short Implant is a flat-bottom Shaping Drill, which is implant specific. It matches the exact shape of the implant without the threads to maximize the initial bone-to-implant contact and implant primary stability.

Short implant placement guidelines

When placing short implants in areas of deficient bone height, following the recommended surgical protocols based on the bone type and using the original instruments and drills is critical to achieve good primary stability of the implants (Fig.4). Moreover, taking certain steps can increase the likelihood of long-term success. The author has developed the following guidelines:

Platform switching: After connection of implants to abutments and exposure to the oral environment, routine loss of approximately 1.5 to 2 mm of vertical bone has long been recognized to occur.⁴⁶ Such changes in the crestal bone can profoundly affect treatment outcomes; the discovery that significantly less peri-implant bone loss occurs when smaller diameter abutments are connected to larger diameter implants⁴⁷ was thus highly significant. Since then, platform switching has become widely accepted as an effective strategy for mitigating post-restorative peri-implant bone loss and increasing overall functional and aesthetic success. Given the fact that short implants are indicated for sites that are vertically deficient to begin with, preventing any additional bone loss is particularly important. When Telleman et al recently examined the impact of platform switching upon peri-




Fig. 5. A 5 mm diameter T3 Short Implant platform switched with a 4.1 mm diameter healing abutment.




Fig. 6. A 6 mm diameter T3 Short Implant with a 4.1 mm diameter healing abutment (double platform switching).

implant bone remodeling around short posterior implants, they found it to be significantly effective.⁴⁸ In all cases, the author thus recommends connecting a smaller diameter abutment to short implants (Figs. 5-6).

Splinting: Splinting of short implant crowns is recommended in order to decrease lateral forces on the prosthesis and reduce stresses on the short implants.⁴⁹ This is true regardless of whether short implants exclusively have been placed or they are being used in combination with standard length implants. When Yilmaz et al compared the strain generated by splinted and non-splinted short implant crowns, they concluded that splinting may provide a more even strain distribution during functional loading.⁵⁰ While it is not possible to splint a single crown supported by a single short implant, an excellent 10-year cumulative survival rate (98.3%) recently was documented for short implants supporting single posterior crowns.⁵¹ Lai et al concluded that a single crown supported by a short implant is a predictable treatment modality. However, as the survival rate for such implants placed in Type IV bone was lower (94%), they cautioned that short implants should be placed in Type IV bone with caution (Figs. 7-8).



Figs. 7, 8. Two T3® Short Implants splinted to a longer implant in a three unit bridge. Note the platform switching on the two T3 Short Implants and the crestal bone preservation one year after implant placement.



Fig. 9. Recommended surgical drilling protocol for crestal placement of 5 mmD x 5 mmL T3 Short Implants. See Surgical Manual ZBINSTMT3S for detailed instructions.

Underpreparation of the osteotomy: The closer contact between an implant and the surrounding bone that results from high insertion torque values (more than 50 Ncm) has been shown to result in more predictable results.⁵² To achieve high insertion torques for short implants placed in Type III and Type IV bone sites, the author recommends underpreparation of the osteotomy following the recommended surgical drilling protocol (Fig. 9).

The crown/implant ratio: Placement of short implants in severely resorbed ridges often increases the crown/implant (C/I) ratio. Some studies have suggested this may lead to greater implant failure rates.³⁰ Some clinicians have considered the greater crown height to be a vertical cantilever that could increase the peri-implant bone

stress⁵³ and eventually result in crestal bone loss, implant failures, or prosthetic complications.⁵⁴⁻⁵⁶ However, recent studies have cast doubt upon these concerns. When Tawil et al followed 262 short, smooth-surfaced implants (for a mean of 53 months), they found no correlation between the C/I ratio or occlusal table and peri-implant bone loss. They concluded that even when the C/I ratio had increased by two to three times, it did not appear to be a biomechanical risk factor if the force orientation and load distribution were favorable. Others have also found that the C/I ratio does not appear to reliably predict implant survival.^{57,58} Although the C/I ratio does not by itself represent a biomechanical risk factor, a very high ratio may lead to mechanical failures such as abutment screw loosening or fracture (Figs. 10-11).

Clinical Guidelines



Fig. 10. Two T3® Short Implants with a high crown/implant ratio are splinted in the same prosthesis to reduce biomechanical stress.



Fig. 11. A T3 Short Implant with an unfavorable crown/implant ratio splinted to a longer implant to distribute loading forces.

Staging of treatment: When short implants were first introduced, use of a staged approach was suggested, leaving the implants submerged to protect the initial phase of osseointegration and avoid the risk of implant failures due to micromovement or contamination.⁵⁹ However, patients often find it uncomfortable to wear removable provisional prostheses during the initial implant-integration phase. The ability to deliver a fixed prosthesis immediately after implant insertion is a major advantage.⁶⁰ Standard length implants placed in selected patients and immediately loaded have been shown to have survival rates comparable to those placed using standard staged procedures even in the presence of poor quality bone, if high insertion torque values (more than 40 Ncm) can be obtained during implant insertion.^{52,61} The author believes the only indication for submerging short implants is an inability to achieve primary stability because of poor bone quality, for example, or inadequate osteotomy site preparation. In all other circumstances, a single-stage approach is preferable. If adequate insertion torque (>50 Ncm) can be achieved for each of the implants, immediate restoration with a healing abutment can be accomplished. When Cannizzaro et al in 2008 compared the outcomes of 7 mm-long implants that were immediately and early loaded, they found survival rates above 96% for both groups after nine months of loading, with no statistically significant differences between the two groups for implant losses, complications, mean marginal bone level changes, and patient preferences⁶² (Figs 12-15).

Implant diameter selection: A minimum of 1 mm to 1.5 mm of bone should be maintained buccal to the implant to avoid buccal soft-tissue recession. Selection of the implant diameter should be based upon this criteria (Figs. 16-18).

Number of implants: In posterior partially edentulous cases, the rule of one implant per tooth should be applied for immediate loading cases. In full-arch cases, it is not necessary due to the cross-arch stabilization obtainable by splinting the provisional restoration (Figs. 19-21).

Connective tissue: An adequate band of keratinized tissue should be present around the implants. The significance of the presence of keratinized mucosa on long-term implant health has been well documented in the literature^{63,64} (Fig. 22).



Figs. 12. Occlusal view of a case performed using flapless, single-stage approach with exposed healing abutments.



Figs. 13. Occlusal view after 4 months of healing.



Figs. 14. Clinical case including a standard length implant in the first premolar and two T3 Short Implants in the second premolar and first molar positions.



Figs. 15. The implants were placed in healed sites in a single stage procedure. A screw-retained bridge out of occlusion, was used as a provisional restoration.



Fig. 16. Occlusal view of the restorative platform of a 4.0 mm D x 11.5 mm L Ex Hex Implant, a 5.0 mm D and a 6.0 mm D T3 Short Implant. The hex size is the same for all three implant diameters.




Fig. 17. Clinical case with a 4.0 mm diameter implant in the premolar site and a 5.0 mm diameter T3 Short Implant in the molar site, allowing for a minimum of 1 mm of buccal bone around both sites.

Clinical Guidelines



Fig. 18. Clinical case with a single 6 mm diameter T3® Short Implant with 1 mm to 2 mm of bone surrounding the implant.



Fig. 19. Posterior partially edentulous case with three T3 Short Implants in place. One implant per tooth was placed as immediate provisionalization was desired.




Fig. 20. Immediate loading of a full mandible with a fixed provisional prosthesis on six implants. The left quadrant posterior implants are two T3 Short Implants.




Fig. 21. Immediate loading of a full mandible with a fixed provisional prosthesis on four implants, with the two posterior implants being T3 Short Implants.



Fig. 22. Connective tissue graft in the buccal side around a T3 Short Implant to achieve an adequate thickness.

Conclusion

The use of short implants makes it possible to provide implant-supported restorations without the need to vertically augment atrophic ridges. The posterior zones can be restored in less time with less risk of complications normally associated with grafting procedures and with less treatment costs.

Francesco Amato, MD, DDS, PhD[†]



Dr. Francesco Amato completed his MD degree at the University of Catania, Italy in 1991. In 1992–1993 he completed a one year full time Advanced Program for International Dentists in Periodontics at New York University College of Dentistry followed by two years full time Advanced Program for International Dentists in Implant Dentistry at New York University College of Dentistry, 1993–1995. He completed his continuing Education Program in Implant Dentistry at New York University College of Dentistry, 1994–1995. He received his Ph.D. Biopharmaceutical Microbiology at the University of Catania, Italy, 1994–1997. He has published numerous articles in international journals; is a lecturer in National and International Conferences and Courses. He is in private practice specializing in Periodontics and Implant Dentistry in Catania, Italy, and is a Clinical Professor in the Master of Periodontology at Universitat Internacional de Catalunya, Spain, a Visiting Professor in the Department of Periodontology at Columbia University, New York, and an International lecturer for the Continuing Dental Education at New York University, New York.

[†]The contributing clinician has a financial relationship with Zimmer Biomet Dental resulting from speaking engagements, consulting engagements, and other retained services.

References

1. Shillingburg HT Jr, Hobo S, Whitsett LE, et al. *Fundamentals of fixed prosthodontics*. Chicago: Quintessence, 1997:89-90.
2. Glantz PO, Nilner K. Biomechanical aspects of prosthetic implant-bone reconstructions. *Periodontol 2000* 1998;17:119-124.
3. Renouard F, Nisand D. Short implants in the severely resorbed maxilla: A 2-year retrospective clinical study. *Clin Implant Dent Relat Res* 2005;6:7 (Suppl 1):S104-S110.
4. Tallgren A. The reduction in face height of edentulous and partially edentulous subjects during long-term denture wear: A longitudinal roentgenographic cephalometric study. *Acta Odontol Scand* 1966;24:195-239.
5. Misch CE. Implant design considerations for the posterior regions of the mouth. *Implant Dent* 1999;8:376-386.
6. Aguilar-Maimban CO. Available bone is the foremost criterion in the insertion of endosteal implants. *J Philipp Dent Assoc* 1996;47:3-21.
7. Nevins M, Langer B. The successful application of osseointegrated implants to the posterior jaw: A long-term retrospective study. *Int J Oral Maxillofac Implants* 1993;8:428-432.
8. Esposito M, Grusovin MG, Felice P, et al. Interventions for replacing missing teeth: horizontal and vertical bone augmentation techniques for dental implant treatment. *Cochrane Database Syst Rev* 2009;(4):CD003607.
9. Chiapasco M, Zaniboni M, Rimondini L. Autogenous onlay bone grafts vs. alveolar distraction osteogenesis for the correction of vertically deficient edentulous ridges: A 2-4 year prospective study on humans. *Clin Oral Implants Res* 2007;18:432-440.
10. Nevins M, Al Hezaimi K, Schupbach P, et al. Vertical ridge augmentation using an equine bone and collagen block infused with recombinant human platelet-derived growth factor-BB: A randomized single-masked histologic study in non-human primates. *J Periodontol* 2012;83:878-884.
11. Chiapasco M, Casentini P, Zaniboni M. Bone augmentation procedures in implant dentistry. *Int J Oral Maxillofac Implants* 2009;24(Suppl):237-259.
12. Das Neves FD, Fones D, Bernardes SR, et al. Short implants – an analysis of longitudinal studies. *Int J Oral Maxillofac Implants* 2006;21:86-93.
13. Feldman S, Boitel N, Weng D, et al. Five-year survival distributions of short-length (10mm or less) machined-surfaced and Osseotite implants. *Clin Implant Dent Relat Res* 2004;6:16-23.
14. Neldam CA, Pinholt EM. State of the art of short dental implants: A systematic review of the literature. *Clin Implant Dent Relat Res* 2012;14:622-632.
15. Hägg D, Deporter DA, Pilliar RM, et al. A targeted review of study outcomes with short (< or =7 mm) endosseous dental implants placed in partially edentulous patients. *J Periodontol* 2004;75:798-804.
16. Van Steenberghe D, Lekholm U, Bolender C, et al. Applicability of osseointegrated oral implants in the rehabilitation of partial edentulism: A prospective multicenter study on 558 fixtures. *Int J Oral Maxillofac Implants*; 1990;5:272-281.
17. Jemt T. Failures and complications in 391 consecutively inserted fixed prostheses supported by Branemark implants in edentulous jaws: a study of treatment from the time of prosthesis placement to the first annual checkup. *Int J Oral Maxillofac Implants* 1991;6:270-276.
18. Friberg B, Jemt T, Lekholm U. Early failure in 4641 consecutively placed Branemark dental implants: A study from stage I to the connection of completed prostheses. *Int J Oral Maxillofac Implants* 1991;6:142-146.
19. Jemt T, Lekholm U. Implant treatment in edentulous maxillae: A 5-year follow-up report on patients with different degrees of jaw resorption. *Int J Oral Maxillofac Implants* 1995;10:303-311.
20. Winkler C, Zarb GA. Treatment outcomes of patients with implant-supported fixed partial prostheses. *Int J Oral Maxillofac Implants* 1998;13:204-211.
21. Lekholm U, Gunne J, Henry P, et al. Survival of the Branemark implant in partially edentulous jaws: A 10-year prospective multicenter study. *Int J Oral Maxillofac Implants* 1999;14:639-645.
22. Bahat O. Branemark system implants in the posterior maxilla: A clinical study of 660 implants followed for 5 to 12 years. *Int J Oral Maxillofac Implants* 2000;15:646-653.
23. Winkler S, Morris HF, Ochi S. Implant survival to 36 months as related to length and diameter. *Ann Periodontol* 2000;5:22-31.
24. Renouard F, Nisand D. Short implants in the severely resorbed maxilla: A 2-year retrospective clinical study. *Clin Implant Dent Relat Res* 2005;7 (Suppl 1):S104-110.
25. Misch CE, Steingra J, Barboza E, et al. Short dental implants in posterior partial edentulism: A multicenter retrospective 6-year case series study. *J Periodontol* 2006;77:1340-1347.
26. Annibali S, Cristalli M, Dell'Aquila D, et al. Short dental implants: A systematic review. *J Dent Res* 2012;91:25-32.
27. das Neves FD, Fones D, Bernardes SR, et al. Short implants: An analysis of longitudinal studies. *Int J Oral Maxillofac Implants* 2006;21:86-93.
28. Al-Hashedi AA, Talyeb Ali TB, Yunus N. Short dental implants: An emerging concept in implant treatment. *Quintessence Int* 2014;45(6):499-514.
29. Misch CE, Steingra J, Barboza E, et al. Short dental implants in posterior partial edentulism: A multicenter retrospective 6-year case series study. *J Periodontol* 2006;77:1340-1347.
30. Wennerberg A, Albrektsson T. Effects of titanium surface topography on bone integration: A systematic review. *Clin Oral Implants Res* 2009;20:172-184.
31. Menchero-Cantalejo E, Barona-Dorado C, Cantero-Alvarez M, et al. Meta-analysis on the survival of short implants. *Med Oral Patol Oral Cir Bucal* 2011;16:e546-e551.
32. Romeo E, Bivio A, Mosca D, et al. The use of short dental implants in clinical practice: literature review. *Minerva Stomatol* 2010;59:23-31.
33. Kotsovilis S, Fourmousis I, Karoussis JK, et al. A systematic review and meta-analysis on the effect of implant length on the survival of rough-surface dental implants. *J Periodontol* 2009;80:1700-1718.
34. Renouard F, Nisand D. Impact of implant length and diameter on survival rates. *Clin Oral Implants Res* 2006;17 Suppl 2:35-51.
35. Arlin ML. Short dental implants as a treatment option: Results from an observational study in a single private practice. *Int J Oral Maxillofac Implants* 2006;21:769-776.
36. Feldman S, Boitel N, Weng D, et al. Five-year survival distributions of short-length (10mm or less) machined-surfaced and Osseotite implants. *Clin Implant Dent Relat Res* 2004;6:16-23.

References

37. Monje A, Fu J-H, Chan H-L, et al. Do implant length and width matter for short dental implants (<10 mm)? A meta-analysis of prospective studies. *J Periodontol* 2013 Dec;84(12):1783-91.
38. Tawil G, Younan R. Clinical evaluation of short, machined-surface implants followed for 12 to 92 months. *Int J Oral Maxillofac Implants* 2003;18:894-901.
39. Anitua E, Tapia R, Luzuriaga F, et al. Influence of implant length, diameter and geometry on stress distribution: A finite element analysis. *Int J Periodontics Restorative Dent* 2010;30:89-95.
40. Lum LB. A biomechanical rationale for the use of short implants. *J Oral Implantol* 1991;17:126-131.
41. Lum LB, Osier JF. Load transfer from endosteal implants to supporting bone: An analysis using statics. Part 1: Horizontal loading. *J Oral Implantol* 1992;18:343-348.
42. Lum LB, Osier JF. Load transfer from endosteal implants to supporting bone: An analysis using statics Part 2: Axial loading. *J Oral Implantol* 1992;18:349-353.
43. Holmgren ET, Seckinger RJ, Kilgren LM, et al. Evaluating parameters of osseointegrated dental implants using finite element analysis: A 2-dimensional comparative study examining the effects of implant diameter, implant shape and load direction. *J Oral Implantol* 1998;24:80-88.
44. Himmlova L, Donstalova T, Kacovsky A, et al. Influence of implant length and diameter on stress distribution: A finite element analysis. *J Prosthet Dent* 2004;91:20-25.
45. Anitua E, Tapia R, Luzuriaga F, et al. Influence of implant length, diameter and geometry on stress distribution: A finite element analysis. *Int J Periodontics Restorative Dent* 2010;30:89-95.
46. Hermann JS, Cochran DL, Nummikoski PV, et al. Crestal bone changes around titanium implants. A radiographic evaluation of unloaded nonsubmerged and submerged implants in the canine mandible. *J Periodontol* 1997;68:1117-1130.
47. Lazzara RJ, Porter SS. Platform switching: A new concept in implant dentistry for controlling postrestorative crestal bone levels. *Int J Periodontics Restorative Dent* 2006;26:9-17.
48. Telleman G, Raghoobar GM, Vissink A, et al. Impact of platform switching on peri-implant bone remodeling around short implants in the posterior region, 1-year results from a split-mouth clinical trial. *Clin Implant Dent Relat Res* 2014;16:70-80.
49. Pierrisnard L, Renouard F, Renault P, et al. Influence of implant length and bicortical anchorage on implant stress distribution. *Clin Implant Dent Relat Res* 2003;5:254-262.
50. Yilmaz B, Seidt JD, McGlumpy EA, et al. Comparison of strains for splinted and nonsplinted screw-retained prostheses on short implants. *Int J Oral Maxillofac Implants* 2011;26:1176-1182.
51. Lai HC, Si MS, Zhuang LF, et al. Long-term outcomes of short dental implants supporting single crowns in the posterior region: A clinical retrospective study of 5 to 10 years. *Clin Oral Implants Res* 2013;24:230-237.
52. Amato F, Polara G. A prospective evaluation of a novel implant designed for immediate loading. *Int J Periodontics Restorative Dent* 2014;34 Suppl 3:s43-s49.
53. Kitamura E, Stegaroiu R, Nomura S, et al. Biomechanical aspects of marginal bone resorption around osseointegrated implants: Considerations based on a three-dimensional finite element analysis. *Clin Oral Implants Res* 2004;15:401-412.
54. Rangert BR, Sullivan RM, Jemt TM. Load factor control for implants in the posterior partially edentulous segment. *Int J Oral Maxillofac Implants* 1997;12:360-370.
55. Misch CE, Suzuki JB, Misch-Dietsh FM, et al. A positive correlation between occlusal trauma and peri-implant bone loss: Literature support. *Implant Dent* 2005;14:108-116.
56. Rangert B, Krogh PH, Langer B, et al. Bending overload and implant fracture: A retrospective clinical analysis. *Int J Oral Maxillofac Implants* 1995;10:326-334.
57. Blanes RJ. To what extent does the crown-implant ratio affect the survival and complications of implant-supported reconstructions? A systematic review. *Clin Oral Implants Res* 2009;20:67-72.
58. Schulte J, Flores AM, Weed M. Crown-to-implant ratios of single tooth implant-supported restorations. *J Prosthet Dent* 2007;98:1-5.
59. Brånemark P-I, Hansson BO, Adell R, et al. Osseointegrated implants in the treatment of the edentulous jaw. Experience from a 10-year period. *Scand J Plast Reconstr Surg Suppl* 1977;16:1-132.
60. Schropp L, Isidor F, Kostopoulos L, Wenzel A. Patient experience of, and satisfaction with, delayed-immediate vs. delayed single-tooth implant placement. *Clin Oral Implants Res* 2004;15:498-503.
61. Ottoni JM, Oliveira ZF, Mansini R, et al. Correlation between placement torque and survival of single-tooth implants. *Int J Oral Maxillofac Implants* 2005;20:769-776.
62. Cannizzaro G, Leone M, Torchio C, et al. Immediate versus early loading of 7-mm-long flapless-placed single implants: A split-mouth randomized controlled clinical trial. *Eur J Oral Implantol* 2008;4:277-292.
63. Lin GH, Chan HL, Wang HL. The significance of keratinized mucosa on implant health: A systematic review. *J Periodontol* 2013;84:1755-1767.
64. Linkevicius T, Puisys A, Steigmann M, et al. Influence of vertical soft tissue thickness on crestal bone changes around implants with platform switching: A comparative clinical study. *Clin Implant Dent Relat Res* 2015;17:6:1228-1236.